



Meter Training Seminar

Arkansas Bureau of Standards

Objectives

- Know and locate all applicable laws that pertain to registered servicepersons
- Seal and Decal Policies
- NTEP Requirement and Certificate of Conformance
- Official Annual Test Reports
- Requirement to adjust device as close as practicable to zero
- Rejected Reports
- Proper Completion of Placed in Service Report (Form 1822)
- What to do if a device is rejected



Laws pertaining to Registered Service Agencies and Persons

Arkansas Title 4 – Business and Commercial Law

- ***AR Code § 4-18-308***
- ***AR Code § 4-18-322*** (Prohibited acts)
- ***AR Code § 4-18-323*** (Civil Penalties)
- ***AR Code § 4-18-324*** (Criminal Penalties)
- ***AR Code § 4-18-344***



AR Code § 4-18-344

- (c) A registered service agent shall perform the recalibration if the inspection or test indicates the bulk meter or liquefied petroleum gas metering device, pump, or scale needs to be recalibrated.
- (d)
 - (1) After the approval of a decal by the Arkansas Bureau of Standards, a registered service agent shall place an approved decal conspicuously on the bulk meter or liquefied petroleum gas metering device, pump, or scale which indicates that it is suitable for trade in accordance with the National Institute of Standards and Technology Handbooks 44 and 112, as adopted by the bureau.
 - (3) A registered technician shall place an approved security seal on the device to prevent any unauthorized access to the adjusting mechanism unless otherwise authorized by the bureau.



AR Code § 4-18-344

- (e) The registered service agent shall provide a copy of all bureau-approved inspection and test reports to the bulk meter or liquefied petroleum gas metering device, pump, or scale owner and to the director.

Note : A copy of the test report **MUST** be left on site at the time of the inspection

- (f)
 - (1) The registered service agent shall retain a copy of all inspection and test reports for a period of three (3) years.
 - (2) The owner of the device shall retain a copy of all inspection and test reports at the device location for a period of three (3) years.



Voluntary Registration

- Officially Place into Service new, used, or rejected devices
- Conduct Calibration (Annual Calibration)
- Remove or Add State Approved Decals
- Authorizes Servicepersons to remove red tags and seals and place rejected devices back into service
- Service person acknowledges that they have in possession and will use all necessary testing equipment and standards; has full knowledge of Arkansas Weight & Measures laws and NIST Handbooks 44, 112, & 130 regulations.
- Director is authorized to suspend or revoke a certificate of registration for:
 - found taking unfair advantage of an owner of a device
 - failure to have test equipment or standards certified: Return to Redo
 - failure to use adequate testing equipment
 - failure to adjust commercial devices to comply with NIST Handbook 44 following a service or repair
 - Repeated Offenses



Voluntary Registration of Servicepersons and Service Agencies NIST Handbook 130

- Section IV. Uniform Regulation
- Subsection D. Uniform Regulation for the Voluntary Registration of Servicepersons and Service Agencies for Commercial Weighing and Measuring Devices



Security Seals

- Devices requiring a physical seal must have one (Red Tag/Rejection)
- Seals must be properly affixed to prevent tampering and adjustments. No loose wire.
- Apply seals only if you need to calibrate a device
- Lead seals must be replaced with state approved seals (Red Tag/Rejection)
- Must be stamped with ARWM one side/Registration number on other side
- NTEP Certificate of Conformance describes the sealing required on each device
- Testing Equipment/Standards Must be Sealed (Red Tag/Rejection)

- The Bureau of Standards DOES NOT provide decals or seals:
 - Decal and seal ordering information found on our website



Security Seals

No Seal



Lead Seal – Must be removed and replaced with state approved seal



Security Seals



Not Sealed



Not Properly Sealed. Remove/Cut Excess Wire



Security Seals



Security Seals

Audit Trail Devices may not require a physical yellow seal



Sealing of Test Measure or Prover

Sealing of test measure or prover is required and subject to being red tagged and rejected if found without seal.



Decals

- Apply current annual (GREEN) decal only after testing and approving the device.
- Each Grade will Require an annual Decal
- Remove all old decals before applying a new decal. No stacking decals.
- Decals shall be affixed as close as possible to the customer indication as possible
- Punch hole in decal for year of test.
- You will be required to return to the facility to correct any issues



Decals

No Stacking Decals



No Yellow Decals Allowed



Decals

Remove all old decals



Decals

No Damaged Decals



Missing Decals



Decals

Punch a hole into year of annual test



Decals

DEF pumps are subject to annual test. Missing decal.



National Type Evaluation Program



Locating the Devices

National Type Evaluation Program (NTEP)

Certificate of Conformance (CC)

- Open Link - <https://www.ncwm.com/ntep-certificates>
- Enter in Devices NTEP # located on data plate:
 - Example NTEP CC No. 02-136



NIST Handbook 130 – Regulation for National Type Evaluation


- Section IV. Uniform Regulation
- Subsection E. Uniform Regulation for National Type Evaluation



Wayne Fueling Systems LLC

Retail Motor Fuel Dispenser / 123/ABCDEF/XX/YY (Generic Name: Orion)

- Certificate Provides example of where to locate the physical security seal



National Type Evaluation Program

- Devices used for trade can have either INACTIVE or ACTIVE status through NCWM
- Any device found to be pre NTEP must obtain an exemption from the director to be used commercially.





OFFICAL ANNUAL GAS STATION TEST REPORTS

Official Annual Test Reports

- When to Perform Annual Tests
- Improper Testing-Adjustments-Calibrations
- Gas Station Test Report and Examples (Report Revised 2023)
- When a report is rejected



When to Perform an Annual Test

- When hired to do by an Arkansas business.
- When installing new or used devices at a location.
- Complete one Official Annual Test Report per year (when placing decals)
-unless told otherwise by our office



Improper Testing

- Tests should be performed in accordance with the prescribed Examination Procedure Outline (EPO) as written in NIST Handbook 112. EPO's No. 21 and No. 22.
- Test should be performed as delivered to the consumer.
 - Pump is operating under normal service conditions
 - NIST procedure is to test to the device, not to the test measure
 - 5 Gallons delivered as shown by the device
- Do not test in stand-alone or calibration mode, this is obvious on reports and will be rejected.
- If a device is adjusted, it must be retested as delivered to the customer.
- Service agencies are subject to having registration revoked as well as civil/criminal penalties if repeated offenses occur.



How many
pumps are in this
picture?

Products?



Prohibited Product Names

Product- Identifying terms that are unacceptable on report:


Examples: Real Gas, Hy-Test, Low-Test, Blended 87, 89, 92, Pump #1, Pump #2, Pump #3, etc. Brand Names

Note: Brand Names are allowed on a case-by-case basis. Contact the Bureau with any questions.



Gas Station Test Report

NTEP CC Number – Must have NTEP CC # on marking information if installed after January 1st 2003





PLANT INDUSTRIES
DIVISION

OFFICIAL ANNUAL TEST REPORT FOR GAS STATIONS

INCORRECT REPORT

Bureau of Standards
4608 West 61st Street
Little Rock, Arkansas 72209
bureau@agriculture.arkansas.gov
501.570.1159

Number of Pumps Adjusted Incorrect

DATE: 01/01/2024		LOCATION NAME: Johns Gas Station #55				EMAIL: john@somewhere.com			PHONE: 501-555-5555		
COUNTY: Garland		PHYSICAL ADDRESS: 123 Main Street				CITY: Hot Springs			STATE: AR		ZIP: 71901
MAILING ADDRESS:		SAME				CITY:			STATE:		ZIP:
NO. PUMPS: 26		NO. PUMPS ADJUSTED: 0		BLEND PUMPS: YES ☒ NO ☒			NO. TANKS BELOW GROUND: 3				
SERVICE AGENCY: John's Testing Service LLC				REG. NO: M0/W0		PHONE: 501-333-333			NO. TANKS ABOVE GROUND: 0		
IDENTIFY PRODUCTS BY CODE, IF OTHER SPECIFY: UNLEADED-U, MIDGRADE-M, PREMIUM-P, DIESEL-D, OFF-ROAD DIESEL-OD *If completing this form electronically select the product tested from the available drop down menu choices below.						RECORD ERROR IN CUBIC INCHES (in ³) + or - AS LEFT			PRICE COMPUTATION = GALLONS DELIVERED X UNIT PRICE		
PRODUCT	PUMP #	MAKE	MODEL	SERIAL #	NTEP CC #	NORMAL TEST	SPECIAL TEST	AS LEFT	GALLONS DELIVERED	UNIT PRICE	PRICE COMPUTATION
Unleaded	1	Wayne	B23	12176D	02-136	-1	-2	-1	10.002	2.599	26.00
Premium						+4	+6	+1	25.003	2.999	74.99
Plus							0	0	5.001	2.799	13.99
Diesel						-1	-2	-1	10.002	3.599	36.00
REMARKS: Pump #3 Out of Order											
By signing this form, the Servicing Agent states; (1) the devices above were inspected and tested in accordance with the current Examination Procedure Outline (EPO) as found in the National Institute of Standards and Technology's Handbook 112 and Handbook 44; (2) the devices above comply with all applicable requirements as specified in Handbook 44.											
John James		John James				John Doe			John Doe		
Service Agent Signature		Service Agent Printed Name				Owner / Operator Signature			Owner / Operator Printed Name		



PLANT INDUSTRIES
DIVISION

OFFICIAL ANNUAL TEST REPORT
FOR
GAS STATIONS

INCORRECT REPORT

Bureau of Standards
4608 West 61st Street
Little Rock, Arkansas 72209
bureau@agriculture.arkansas.gov
501.570.1159

Gallons Delivered Error

DATE: 01/01/2024		LOCATION NAME: Johns Gas Station #55				EMAIL: john@somewhere.com			PHONE: 501-555-5555		
COUNTY: Garland		PHYSICAL ADDRESS: 123 Main Street				CITY: Hot Springs			STATE: AR		ZIP: 71901
MAILING ADDRESS:		SAME				CITY:			STATE:		ZIP:
NO. PUMPS: 26		NO. PUMPS ADJUSTED: 1		BLEND PUMPS: YES ☒ NO ☒			NO. TANKS BELOW GROUND: 3				
SERVICE AGENCY: John's Testing Service LLC				REG. NO: M0/W0		PHONE: 501-333-333		NO. TANKS ABOVE GROUND: 0			
IDENTIFY PRODUCTS BY CODE, IF OTHER SPECIFY: UNLEADED-U, MIDGRADE-M, PREMIUM-P, DIESEL-D, OFF-ROAD DIESEL-OD *If completing this form electronically select the product tested from the available drop down menu choices below.						RECORD ERROR IN CUBIC INCHES (in ³) + or - AS LEFT			PRICE COMPUTATION = GALLONS DELIVERED X UNIT PRICE		
PRODUCT	PUMP #	MAKE	MODEL	SERIAL #	NTEP CC #	NORMAL TEST	SPECIAL TEST	AS LEFT	GALLONS DELIVERED	UNIT PRICE	PRICE COMPUTATION
Unleaded	1	Wayne	B23	12176D	02-136	-1	-2	-1	10.002	2.599	26.00
Premium						+4	+6	+1	10.003	2.999	74.99
Plus							0	0	5.001	2.799	13.99
Diesel						-1	-2	-1	10.002	3.599	36.00
REMARKS: Pump #3 Out of Order											
By signing this form, the Servicing Agent states; (1) the devices above were inspected and tested in accordance with the current Examination Procedure Outline (EPO) as found in the National Institute of Standards and Technology's Handbook 112 and Handbook 44; (2) the devices above comply with all applicable requirements as specified in Handbook 44.											
John James		John James				John Doe			John Doe		
Service Agent Signature		Service Agent Printed Name				Owner / Operator Signature			Owner / Operator Printed Name		



PLANT INDUSTRIES
DIVISION

OFFICIAL ANNUAL TEST REPORT FOR GAS STATIONS

INCORRECT REPORT

Bureau of Standards
4608 West 61st Street
Little Rock, Arkansas 72209
bureau@agriculture.arkansas.gov
501.570.1159

Incorrect Test Procedure – AS LEFT

DATE: 01/01/2024		LOCATION NAME: Johns Gas Station #55				EMAIL: john@somewhere.com			PHONE: 501-555-5555		
COUNTY: Garland		PHYSICAL ADDRESS: 123 Main Street				CITY: Hot Springs			STATE: AR		ZIP: 71901
MAILING ADDRESS:		SAME				CITY:			STATE:		ZIP:
NO. PUMPS: 26		NO. PUMPS ADJUSTED: 26		BLEND PUMPS: YES ☒ NO ☒			NO. TANKS BELOW GROUND: 3				
SERVICE AGENCY: John's Testing Service LLC				REG. NO: M0/W0		PHONE: 501-333-333			NO. TANKS ABOVE GROUND: 0		
IDENTIFY PRODUCTS BY CODE, IF OTHER SPECIFY: UNLEADED-U, MIDGRADE-M, PREMIUM-P, DIESEL-D, OFF-ROAD DIESEL-OD *If completing this form electronically select the product tested from the available drop down menu choices below.						RECORD ERROR IN CUBIC INCHES (in ³) + or - AS LEFT			PRICE COMPUTATION = GALLONS DELIVERED X UNIT PRICE		
PRODUCT	PUMP #	MAKE	MODEL	SERIAL #	NTEP CC #	NORMAL TEST	SPECIAL TEST	AS LEFT	GALLONS DELIVERED	UNIT PRICE	PRICE COMPUTATION
Unleaded	1	Wayne	B23	12176D	02-136	-1	-2	0	10.002	2.599	26.00
Premium						+4	+6	0	15.003	2.999	74.99
Plus							0	0	5.001	2.799	27.99
Diesel						-1	-2	0	10.002	3.599	36.00
								0			
								0			
								0			
								0			
								0			
								0			
								0			
								0			
								0			
REMARKS: Pump #3 Out of Order											
By signing this form, the Servicing Agent states; (1) the devices above were inspected and tested in accordance with the current Examination Procedure Outline (EPO) as found in the National Institute of Standards and Technology's Handbook 112 and Handbook 44; (2) the devices above comply with all applicable requirements as specified in Handbook 44.											
John James			John James			John Doe			John Doe		
Service Agent Signature			Service Agent Printed Name			Owner / Operator Signature			Owner / Operator Printed Name		

A Report is Rejected if...

- The test results indicate the possibility the devices were not tested in accordance with the applicable Handbook 112 Examination Procedures Outline (EPO)
- Not Legible
- Not Complete
- Not Correct
- Purpose of the annual test is to ensure the business is compliance with all applicable state laws and regulations.





Bureau of Standards
4608 West 61st Street
Little Rock, Arkansas 72209
bureau@agriculture.arkansas.gov
501.570.1159

Form G-1A Revised 4-21-2023

Acknowledging Compliance



PLANT INDUSTRIES
DIVISION

OFFICIAL ANNUAL TEST REPORT FOR GAS STATIONS

Bureau of Standards
4608 West 61st Street
Little Rock, Arkansas 72209
bureau@agriculture.arkansas.gov
501.570.1159

DATE:		LOCATION NAME:				EMAIL:			PHONE:		
COUNTY:		PHYSICAL ADDRESS:				CITY:			STATE:		ZIP:
MAILING ADDRESS:						CITY:			STATE:		ZIP:
NO. PUMPS:		NO. PUMPS ADJUSTED:		BLEND PUMPS: YES ☐ NO ☐			NO. TANKS BELOW GROUND:				
SERVICE AGENCY:				REG. NO:		PHONE:			NO. TANKS ABOVE GROUND:		
IDENTIFY PRODUCTS BY CODE, IF OTHER SPECIFY: UNLEADED-U, MIDGRADE-M, PREMIUM-P, DIESEL-D, OFF-ROAD DIESEL-OD *If completing this form electronically select the product tested from the available drop down menu choices below.						RECORD ERROR IN CUBIC INCHES (in ³) + or - AS LEFT			PRICE COMPUTATION = GALLONS DELIVERED X UNIT PRICE		
PRODUCT	PUMP #	MAKE	MODEL	SERIAL #	NTEP CC #	NORMAL TEST	SPECIAL TEST	AS LEFT	GALLONS DELIVERED	UNIT PRICE	PRICE COMPUTATION
REMARKS:											
By signing this form, the Servicing Agent states: (1) the devices above were inspected and tested in accordance with the current Examination Procedure Outline (EPO) as found in the National Institute of Standards and Technology's Handbook 112 and Handbook 44; (2) the devices above comply with all applicable requirements as specified in Handbook 44.											
Service Agent Signature				Service Agent Printed Name		Owner / Operator Signature			Owner / Operator Printed Name		

Form G-1A Revised 4-21-2023



TEST REPORT MUST BE LEFT ON SITE AFTER APPROVAL

AR Code § 4-18-344 (c)

-Two Weeks to Turn In

AR Code § 4-18-344 (f)(1)&(2) - 3 Years





Place in Service Reports (Form 1822)


When to complete a Form 1822

- A device being placed into service for the first time. Ex. EVSE
- When a new or used device is being installed at a location in the state
- Any device is that is moved and reinstalled anywhere in the state
- When a device has been rejected by the state



When to complete Form 1822

- If a device is restored to service after official rejection by the Bureau of Standards (Red Tagged)
- Form shall be emailed to the Bureau of Standards within 24 hours of a device being returned to service
- Email: bureau@agriculture.arkansas.gov

A red rectangular tag with a circular hole at the top center. It contains text in black and red. At the top left is a circular logo with 'AR' and 'ARKANSAS DEPARTMENT OF AGRICULTURE' around it. To the right of the logo, it says 'FORM NO 5' and 'This device has been REJECTED'. Below that, it says 'by the Arkansas Department of Agriculture Bureau of Standards.' and 'DO NOT USE THIS DEVICE'. Further down, it says 'This tag must not be removed unless authorized by the Director of Bureau of Standards, 4608 W. 61st, Little Rock, AR 72209'. At the bottom, there are fields for 'Date:', 'Inspector:', 'Product:', 'Serial No.', 'Make:', and 'Capacity:', followed by 'Tag Number:' and 'Nº 419'.

FORM1822.arkansas.gov

Rejection Red Tags and Seals

If a device has been officially rejected:

- Servicepersons are authorized to remove Bureau applied tags and seals to repair devices
- If device is not restored to FULL compliance (HB 44 Requirements)
 - Device is NOT to be returned to service
 - Serviceperson shall reattach the original Bureau applied red tag with serviceperson's properly attached security seal





PLANT INDUSTRIES
DIVISION

OFFICIAL PLACED IN SERVICE REPORT FOR METERS

Bureau of Standards
4608 West 61st Street
Little Rock, Arkansas 72209
burstdm@agriculture.arkansas.gov
501.570.1159

Officially Rejected
Device ☒

Newly Installed
Device ☐

Used Device at a
New Location ☐

Device
Calibration ☐

Name of Owner: John Johnson Phone: (501)-555-5555
Name of Facility: Johns Test Site Phone: (501)-555-3333
Email Address: john@somewhere.com
Mailing Address: PO Box 123 or SAME Hot Springs AR 71901
Physical Address: 123 Main Street Hot Springs Garland 71913

TYPE OF METER	METER DESIGN	STORAGE TANK
Truck Meter ☐	Make: Wayne	☒ Below Ground
Dock Meter ☐	Model: B23	☐ Above Ground
Diesel Pump ☐	Serial #: 12176D	
Gas Pump ☒	Flow Rate: 12gpm	
Mass Flow Meter ☐	National Type Evaluation Program (NTEP) Certificate of Conformance Number: 02-136	
Agri-Meter ☐		
Other Meter ☐		
(Specify):	*(Serial numbers of the same Make and Model may be recorded on the reverse side.)	

This Official Placed in Service Report, and the appropriate Official test report, must be mailed within 24 hours to the Bureau of Standards executed by a Serviceperson representing a Registered Service Agency or a Registered Self-employed Serviceperson for each device calibrated, used device reinstalled at a new location, newly installed device, or an officially rejected device restored to service. You must include the original rejection tag.

This is to certify that I have repaired and/or installed and left as correct in accordance with the current version of the National Institute of Standards and Technology (NIST) Handbook 44, the equipment described above.

Service Agency: Jacks Meter Service LLC Registration #: M0/W0
Serviceperson: Jack Dirt Date of Service: 01/01/2024
Signature: Jack Dirt
Remarks: Owner John Johnson John Johnson





Meter Inspection and Testing Training

Arkansas Bureau of Standards

Objectives

- Upon completion of this presentation, you will be able to understand and describe Inspection for *correctness* & Testing for *accuracy*
 - Referencing Handbook 44 using Handbook 112 (EP0 21&22)
 - General Code Requirements
 - Safety
 - Indicating and Recording Element
 - Marking Requirements
 - Discharge Hose-retail
 - Totalizer
 - Filters (NIST Handbook 130)
 - Testing
 - Sealing



Handbook 112

Examination Procedure Outline (EPO) for Retail Motor-Fuel Dispensers (RMFDs)

EPO No. 21 Non-Blended Product

EPO No. 22 Blended Product

Handbook 44

Device Requirements found in Two Sections

Section 1.10. General Code

Section 3.30. Liquid-Measuring Devices



Handbook 44 Key

- Handbook Key
- Any G- prefix
- General code – Applies to all devices in HB 44
 - G-A – Application
 - G-S – Specifications
 - G-N – Notes
 - G-T – Tolerances
 - G-UR – User Requirements



Inspection:

Safety First !!!

Check the inspection site carefully for safety hazards and take appropriate precautions.

Learn the nature of hazardous products used at or near the inspection site – obtain and read copies of MSDS's.

Know emergency procedures and location and operation of fire extinguisher and emergency shut-offs.

Post safety cones/warning signs and be aware of vehicular and pedestrian traffic patterns.

Use caution in moving in wet, slippery areas.

Use personal protection equipment and clothing appropriate for the inspection site.

Open both sides of dispenser to allow fumes to dissipate before proceeding with the inspection of the dispenser.

If leaks, spills, or exposed wiring cause hazardous testing conditions it is recommended that the testing be discontinued until the unsafe conditions are corrected.

Be sure that a first aid kit is available and that it is appropriate for the type of inspection activity.

H-44 General Code and Liquid-Measuring Devices Code References

1. General considerations.	
Selection.....	G-S.3., G-UR.1.1., G-UR.1.2., G-UR.1.3.
Installation.....	G-S.2, G-UR.2.1., G-UR.2.2., UR.2.1., UR.2.4.
Position of equipment.....	G-UR.3.3.
Accessibility.....	G-UR.2.3.
Assistance.....	G-UR.4.4.
Use and maintenance.....	G-UR.3.1., G-UR.4.1., G-UR.4.2., UR.3.5.
2. Indicating and recording elements.	
Design.....	S.1.1.
Units.....	S.1.2.1., S.1.2.3.(a)
Readability.....	G-S.5., G-S.6. (1/1/77), G-S.7., S.1.4., S.1.5.
Values of intervals.....	G-S.5.3., G-S.5.3.1.
Indication of delivery.....	S.1.6.1.
Money-value divisions	
Analog.....	S.1.6.5.1.
Digital.....	S.1.6.5.2.
Auxiliary indications.....	S.1.6.5.3. (1/1/85)
Unit Price and product identity.....	S.1.6.4.1 (a), S.1.6.4.2., UR.3.2.

Handbook 44 Key

- Handbook Key
- Any code with no G prefix
- Applies to only –
Liquid-Measuring Devices (Section 3.30)
 - A. – Application
 - S. – Specifications
 - N. – Notes
 - T. – Tolerances
 - UR. – User Requirements

Inspection:

Safety First !!!

Check the inspection site carefully for safety hazards and take appropriate precautions.
Learn the nature of hazardous products used at or near the inspection site – obtain and read copies of MSDS's.
Know emergency procedures and location and operation of fire extinguisher and emergency shut-offs.
Post safety cones/warning signs and be aware of vehicular and pedestrian traffic patterns.
Use caution in moving in wet, slippery areas.
Use personal protection equipment and clothing appropriate for the inspection site.
Open both sides of dispenser to allow fumes to dissipate before proceeding with the inspection of the dispenser.
If leaks, spills, or exposed wiring cause hazardous testing conditions it is recommended that the testing be discontinued until the unsafe conditions are corrected.
Be sure that a first aid kit is available and that it is appropriate for the type of inspection activity.

11-44 General Code and Liquid-Measuring Devices Code References

1. General considerations.	
Selection.....	G-S.3., G-UR.1.1., G-UR.1.2., G-UR.1.3.
Installation.....	G-S.2, G-UR.2.1., G-UR.2.2., UR.2.1., UR.2.4., G-UR.3.3.
Position of equipment.....	G-UR.2.3.
Accessibility.....	G-UR.4.4.
Assistance.....	G-UR.3.1., G-UR.4.1., G-UR.4.2., UR.3.5.
Use and maintenance.....	
2. Indicating and recording elements.	
Design.....	S.1.1.
Units.....	S.1.2.1., S.1.2.2., (a)
Readability.....	G-S.5., G-S.6. (1/1/77), G-S.7., S.1.4., S.1.5., G-S.5.3., G-S.5.3.1.
Values of intervals.....	S.1.6.1.
Indication of delivery.....	
Money-value divisions	
Analog.....	S.1.6.5.1.
Digital.....	S.1.6.5.2.
Auxiliary indications.....	S.1.6.5.3. (1/1/85)
Unit Price and product identity.....	S.1.6.4.1 (a), S.1.6.4.2., UR.3.2.



General Code Definitions

- Retroactive Requirements
 - Enforceable on ALL EQUIPMENT
 - Printed in Upright Roman Type in Handbook 44
 - Example – Manufacturer marking requirement.
- Nonretroactive Requirements
 - Printed in *Italics Type* in Handbook 44
 - Enforceable on or after effective date.
 - Example – NTEP CC # marking requirement.





EPO No. 21 & No. 22

Safety

Safety

- Ensure grounded contact of
 - Test Measure to Funnel to Storage
 - Prover Line to Storage
 - Nozzle/Handle to Test Measure
- Avoid Wearing Synthetic Clothing
- Steel/Composite Boots
- Avoid "Switch Loading"
- **DO NOT USE A PLASTIC CONE AS A FUNNEL**



Fuel in Cabinet

Active Fuel Leak - DO NOT LEAVE A DISPENSER LIKE THIS
Fire Marshall & DEQ



Active Fuel Leak

Handle is leaking
when hose is
pressurized

Fire Marshall
DEQ



Safety Hazard

Shear Valve Pried
Open

Fire Marshall
DEQ





EPO No. 21 & No. 22

General Consideration Requirements

General Consideration Requirements

- **Requirements that Must be observed when**
 - Purchasing, Installing, and Inspecting Devices on behalf of device owner
- **Selection**
 - **G-UR.1.1.** - Selection of Equipment – Is the device suitable for the service to which it will be used. Rate of flow, character number and size and location of its indicating and recording elements. DEF Tubs with Non NTEP meters
 - **G-UR.1.2.** - Environment – Equipment should be suitable for the environment its to be installed in. Wind, Weather RFI



General Consideration Requirements

- **Installation Requirements**

- **G-S.2.** - Facilitation of fraud - All equipment attached to shall be design to not facilitate fraud.
- **G-UR.2.1., G-UR.2.2., & UR.2.1.** - Installation of devices, indicating and recording elements shall be installed according to manufacturer instructions. Shall be secure and rigid.
- **UR.2.4.** - Diversion of liquid flow for trucks must have deterrents installed to prevent other vehicles other than the receiving vehicle from getting fuel.

- **Position of Equipment**

- **G-UR.3.3.** - Indication shall be observed from some reasonable position.



Fraudulent/Unapproved Equipment



Magnet




3rd Party WIFI Router



Fraudulent/Unapproved Equipment



Skimmers



Pulser Manipulation



General Consideration Requirements

- **G-UR.2.3.** Accessibility - Owners shall provide access for inspecting, testing, sealing of a device.
- **G-UR.4.4.** Assistance in Testing Operations - Must be supplied by owner or operator
 - Special Equipment -Provide Keys, remote, keypads, tools to enter cabinets
 - Accessories
 - Abnormal Labor (Marinas)



General Consideration Requirements

- **Use of Device**
 - **G-UR.3.1. Method of Operation**
 - Only be operated in a manner that is obviously indicated by construction or indicated by instructions on the equipment
 - **UR.3.5. Steps after Dispensing** – After the delivery to a customer
 - The starting lever shall be returned to shutoff position and zero setback interlock engaged
 - The discharge nozzle returned to designed hanging position
 - Unless indicating and recording elements, if equipped and activated to record, have been returned to a definite zero



General Consideration Requirements

Maintenance Requirements

- **G-UR.4.2. Abnormal Performance**
 - Unstable Indications
 - Abnormal Equipment Performance
 - Examples - Abnormal flow rates (slow), Unusual noises during operation, Vibrations or shaking



Official Rejection for Abnormal Performance: Flow Rate too Slow

Performing Normal Test
Found Releasing ~3GPM

Max Flow Rate 12 GPM
Min Flow Rate 0.5 GPM

Rejected for dispensing too
slow. Consumer Complaint.

Press Play ->



General Consideration Requirements

Maintenance Requirements- Predominance

- **G-UR.4.1. Maintenance of Equipment**

- Equipment in service at a single place of business shall not be considered “maintained in a proper operating condition” if
 - Predominantly, equipment ***of all types or applications*** are found to be in error or direction favorable to the device user (ex. Scales and Pumps)
 - Predominantly, equipment ***of the same type*** or application is found to be in error in a direction favorable to the device user (ex. Pumps)

These sites were found to have errors favorable to one user over another.

- Devices must be maintained in proper operating condition and adjusted as close to zero as possible. There must not be a bias in one direction that is favorable to the device owner
- Sites found to have predominance are subject to entire retest



General Consideration Requirements


Maintenance Requirements – Any attached equipment.

- **G-UR.4.1. Maintenance of Equipment**

- All equipment in service and all mechanisms and devices attached thereto or used in connection therewith shall be continuously maintained in proper operating condition throughout the period of such service.
- Most common violation for devices.
- The following slides provide examples of devices not properly maintained...



Official Rejection for Maintenance of Equipment: Leaking



Official Rejection for Maintenance of Equipment: Leaking

Leak from at or near Breakaway



Leak from Nozzle, Swivel, or at thread




Official Rejection for Maintenance of Equipment: Leaking

Leaking from filter



Leak from meter



Check Bottom of Filter

Official Rejection for Maintenance of Equipment: Leaking from filter

Press Play



Official Rejection for Maintenance of Equipment: Leaking from Meter



Press Play ->

Official Rejection for Maintenance of Equipment: Display not Working



Official Rejection for Maintenance of Equipment: Price Per Gallon Display not Working



Official Rejection for Maintenance of Equipment: Displays not readable

Sun Damage



Sun Damage



Official Rejection for Maintenance of Equipment: Lights Out in ANY Displays


Light Out in PPG Display



Light Out in PPG Display




Official Rejection for Maintenance of Equipment: Segments missing from digital information on Indicators/Displays



Official Rejection for Maintenance of Equipment: Hose Damage

Bulging/Cracking



Official Rejection for Maintenance of Equipment: Hose Damage

Split



Cracking



Official Rejection for Maintenance of Equipment: Damage to Product Identity



Official Rejection for Maintenance of Equipment: Damage to Octane Rating





EPO No. 21 & No. 22

Indicating and Recording Element Requirement

Indicating and Recording Elements

- **Readability Requirements**

- **G-S.7. Lettering**

- All required markings and instructions shall be distinct and easily readable and shall be of such character that they will not tend to become obliterated or illegible.



All lettering associated with Indicators and Displays shall be readable and not become easily obliterated

Example of Legible Lettering. Not easily obliterated.



\All lettering associated with Indicators and Displays shall be readable and not become easily obliterated

Product Identity Not Readable. Cannot be covered.



Indicating and Recording Elements

- Money Value Division

- S.1.6.5.1 Analog Devices—

See the MAV (Tolerance) Table for computed sales when conducting field tests on analog devices. MAV is based on the unit price at the location.

Example – The Unit Price is \$2.50 a gallon at a site. Then the allowable error in total sale can be (+/–) 2 cents when conducting field tests.

Table 1. Money-Value Divisions and Maximum Allowable Variations for Money-Value Computations on Mechanical Analog Computers				
Unit Price		Money-Value Division	Maximum Allowable Variation	
From	To and Including		Design Test	Field Test
0	\$0.25/liter or \$1.00/gallon	1¢	± 1¢	± 1¢
\$0.25/liter or \$1.00/gallon	\$0.75/liter or \$3.00/gallon	1¢ or 2¢	± 1¢	± 2¢
\$0.75/liter or \$3.00/gallon	\$2.50/liter or \$10.00/gallon	1¢ or 2¢	± 1¢	± 2¢
		5¢	± 2½¢	± 5¢



Indicating and Recording Elements

- **Money Value Divisions**
 - **S.1.6.5.3** Auxiliary Elements – AKA Secondary Indicators
 - If a system is equipped with auxiliary indications (Secondary Indicator at POS), all indicated money-value divisions of the auxiliary element shall be identical to those in the primary element (Pump)
 - Simply put, the total sale and unit price at the primary indicator (pump) shall be identical to the secondary indicator at the point-of-sale
 - Common complaint regarding analog devices




Indicating and Recording Elements

- **UR.3.1. Return of Indicating and Recording Elements to Zero**
 - On any dispenser in making retail deliveries, the primary indicating element, and recording element if so equipped, shall be returned to zero before each delivery.
 - Exception – totalizers on key lock operated or other self-operated dispensers



Official Rejection for Primary Indicator not Readily Returning to Zero

Total Sale did not start at Zero. Regulation requires devices to start at zero.



Primary Indicator Readily Returned to Zero



Primary Indicator Readily Returned to Zero



Indicating and Recording Elements

- **S.1.6.7 Recorded Representations**

- A receipt providing the following information shall be available through ***built-in*** or ***separate recording element*** for all transactions conducted with a Point-of-Sale or devices activated by debit cards, credit cards, and/or cash:

1. Total Volume
2. Unit Price
3. Total Computed Price
4. Product identified by - Name, Symbol, Abbreviation, Code Number
5. Dispenser Designation – Alphabetical or Numerical



Required Information on Recorded Representation

5 Requirements

1. Total Volume
2. Unit Price
3. Total Computed Price
4. Product identified by - Name, Symbol, Abbreviation, Code Number
5. Dispenser Designation - Alphabetical or Numerical

Jiffy Trip 0704
550 W. US 412 HWY
SILOAM SPRINGS
AR, 72761
DEALER #: 21

8/13/2024
1:38:12 PM

Pump # 20 - Self
Grade: E10 19.293g
Price/Gal 2.899
Fuel Ttl \$55.93

Completion
WEX Acct: 5102
Swiped
AUTH: 00-428746
Batch: 78 Seq: 11
VEHICLE: 08342
ODOMETER: 67942
INVOICE: 133513
TRAN: 582932

Tell us about
your visit for a
chance to win
a gas gift card!
Gasfeedback.com





EPO No. 21 & No. 22 Marking Requirements

Marking Requirements

- **G-S.1. Identification shall be clearly and permanently marked with -**
 - (a) Manufacturer by name, initials, trademark
 - (b) Model identifier
 - (c) Nonrepetitive Serial Number
 - (d) Current Software Version
 - (e) NTEPCC #



Official Rejection: Required Marking Information Not Readable



Marking Requirements

- **S.4.4.2. Location of Marking Information Shall:**
 - Appear within 24 inches to 60 inches from the base of device
 - Internally or externally provided its permanent and easily readable
 - On a portion that is not easily removed or interchangeable
 - Servicepersons can create nameplates for devices that have missing or damaged marking information. So long as the data plate is accurate and correct.
- *Nonretroactive January 1, 2003*



Required Marking Information Easily Readable



**GILBARCO
VEEDER-ROOT**

Mfd By Gilbarco Inc.
Greensboro, NC USA

ENCORE 7005 RE

SERIAL KEEN397363
NUMBER

MODEL NN1
NUMBER

NTEP CC NO. 02-019

For use with equipment specified in installation instructions.



POWER OPERATED DISPENSING DEVICE FOR FLAMMABLE LIQUIDS
FOR USE IN CLASS I, DIVISION 2 GROUP D HAZARDOUS
LOCATIONS 34GL

LISTED

SERVICE STATION HOSE NOZZLE VALVE
FOR USE ONLY WITH UL LISTED INTERCHANGEPE

CAUTION - Hazard of electrical shock - always disconnect before servicing.
Danger - Lequel de choc électrique - Plus d'un débrancher peut être nécessaire afin de mettre l'appareil hors tension pour réparation.
WARNING - Do not disconnect components, fuses, lamps, etc. while circuit is live.
Attention - Ne pas débrancher les composants, fusibles, supports de lampes, etc. pendant que le circuit est sous tension.

UNIT
Voltage: 115
Hertz: 50/60
Amps: 13.0

Options:

SPEAKER
PUSH TO START
TOTALIZERS
CIN CENTER
INTERCOM



Marking Requirements

- **G-UR.3.4 Responsibility, Money Operated Device** - where employees are not present must clearly and conspicuously display information detailing method for return of monies paid if product or service is not obtained. ie...Name, address and phone number of the local responsible party for the device.
- **S.4.4.1. Discharge Rates must be marked on the device if:**
 - The devices discharge rate is 30 gallons per minute or greater (high flow devices). Maximum and minimum discharge rate shall be marked in accordance with **S.4.4.2.** - 2-5ft from base.
 - Marked Minimum Discharge Rate shall not exceed 20% of marked Maximum Discharge Rate.
 - *[Nonretroactive as of January 1, 1985]*





EPO No. 21 & No. 22

Discharge Hose-Retail Requirements

Discharge Hose Requirements

- **S.3.1. Diversion of Measured Liquid** – There should be no way to divert measured liquid from measuring chamber or the discharge line other than how the device was initially designed.
 - However, two or more delivery **outlets** are permitted if there are means to ensure:
 - Liquid can flow from only one such outlet at a time (multiproduct dispensers)
 - Direction of flow is clearly indicated
- **S.3.2. Exceptions to the diversion of measured liquid include:**
 - Truck refueling devices. Satellite pumps. So long as deterrents are installed to prevent other vehicles, other than the receiving vehicle, from getting fuel.



Discharge Hose Requirements

- **S.3.3. Pump -Discharge Unit** - On a pump-discharge unit (RMFD), equipped with a discharge hose, the discharge hose shall be of the wet-hose type.
- **S.3.5. Discharge Hose, Reinforcement** - Discharge hose shall be reinforced so that the performance (measurement) is not affected by expansion or contraction
- **S.3.6. Discharge Valve (Nozzle)** - Discharge valves (nozzles) may be installed in discharge lines only if the device is of the wet-hose type.
- **S.3.7. Anti-drain Means** - In a wet-hose, pressure-type device, means shall be incorporated to prevent the drainage of the discharge hose.



Discharge Hose Requirements

- **UR.1.1. Discharge Hose.**

UR.1.1.1. Length- the length of a discharge hose on an RMFD hose;

(a) Shall be measure from its housing or outlet of the discharge line to the inlet of the discharge nozzle;

(b) Shall be measured with the hose fully extended if it is coiled, retained, or connected inside a housing and;

(c) Shall not exceed 18 feet unless a need is demonstrated



Discharge Hose Requirements

- **UR.1.1.2. Marinas and Airports**

- **UR.1.1.2.1. Length**

- Length shall be short as practicable, shall not exceed 50 feet unless a need is demonstrated

- **UR.1.1.2.2. Protection**

- Hoses exceeding 26 feet shall be adequately protected from weather and other environmental factors when not in use





EPO No. 21 & No. 22 Totalizers

Totalizers

- **S.5. Totalizers for Retail Dispensers**

- Shall be equipped with a non-resettable totalizer for the quantity delivered through the metering device
- *Nonretroactive January 1, 1995*





NIST Handbook 130 Filter Requirements

Filters

- NIST Handbook 130 – Section IV. (F)
- 4.3. Dispenser Filters
- 4.3.1. Engine Fuel Dispensers.
 - All gasoline, gasoline-alcohol blends, gasoline-ether blends, ethanol fuel, and M85 methanol dispensers shall have a 10 micron or smaller nominal pore-sized filter.
 - All biodiesel, biodiesel blends, diesel, and kerosene dispensers shall have a 30 micron or smaller pore-sized filter.
 - RMFDs **MUST** have filters and **MUST** have the correct size.



Official Rejection for No Filters. RMFDs MUST HAVE FILTERS

No Filter Inside Cabinet



Same Device: No Filter Outside Cabinet



Official Rejection for Incorrect Filter Size

30 Micron Filter on Gasoline Dispenser



Only up to 10-micron pore-sized filters allowed on gasoline dispensers



Official Rejection for Non-Compliant Filter Types

Example of a device that has only mesh wire type cloth basket – Micron Pore Size is 120



Example of a device where the filter can only be accessed by shutting down the pump and by use of a tool. These filters do not have micron sized stenciled or labeled on part



In both examples BOS office will require retrofitting of filters to ensure the business is compliant. Ex. Adding in-Line filters.





EPO No. 21 & No. 22

Pretest Determinations and Notes

Pretest Determinations and Notes

Test Draft Size

- The size of a test draft will depend on the devices Designed Maximum Discharge (Flow) Rate
- If the flow rate is Less than 20 GPM
 - Tests shall include drafts of one or more amounts, including a test draft of at least 5 gallons
- If the flow rate is Greater than or equal to 20 GPM
 - Tests shall include drafts of one or more amounts, including a test draft of at least the amount delivered in one minute (gpm) at the max flow rate of the installation.
 - Example – if listed maximum flow rate 60 GPM, test draft must be at least 60 gallons, preferably larger (100 gallon). Considering if the device under registers.



Pretest Determinations and Notes


Test Measures & Provers

- Reading the Test Standard
 - Standard must be level
 - Leveled by “plumbing the neck”
 - Allow all foam to settle before taking measurement - Diesel
 - Read measurement before fuel evaporates - Gasoline
 - Read the meniscus at eye level.
 - Avoid parallax – Taking readings at different positions from different points



Pretest Determinations and Notes

Reading the Meniscus



Pretest Determinations and Notes

Adjustments/Calibration -

“The Registered Serviceperson or Service Agency is responsible for **installing, repairing, and adjusting** devices such that the devices are adjusted as closely as practicable to zero error.”

Use the applicable tolerances to confirm device compliance after adjusting/calibrating.



Pretest Determinations and Notes

To prepare for testing, determine what tolerance is applicable to the device.

1.10. General Code, G-T. Tolerances

Acceptance Tolerance G-T.1. These tolerances shall apply to equipment;	Maintenance Tolerance G-T.2.
To be put into commercial use for the first time	These tolerances shall apply to equipment in use, except as provided in G-T.1. Acceptance Tolerances
That has been placed in commercial service within the preceding 30 days and is being officially tested for the first time	
That has been returned to commercial service following official rejection for failure to conform to performance requirements and is being officially tested for the first time within 30 days after corrective service	
Being officially tested for the first time within 30 days after major reconditioning or overhaul and	
Undergoing type evaluation	
Undergoing Annual Test and within 30 days of the test. (Office Policy)	



Pretest Determinations and Notes

Tolerance Values Continued...

- **T.2. Tolerance Values** – The RMFD specific applicable maintenance, acceptance, and special test tolerances shall be shown in the NIST Handbook 44, Section 3.30 Liquid-Measuring Device, Table T.2.



Tolerance Values

Table T.2.

1st Row in Table -

Applies to High Flow meters. Meters that deliver greater than 30gpm will have the following tolerances:

Acceptance Tolerances – 0.2%

Maintenance Tolerances – 0.3%

Special Test Tolerances – 0.5%

% in allowable error is based on test draft size

Example – a meter with a max flow rate of 30 gpm or greater can have an acceptance tolerance of + or – 0.2%. If the test draft is 100 gallons, the tolerance is 0.2 gallons.

0.2 gallons = 46.2 cubic inches

Coverison: 231 cubic inches per 1 gallon



Accuracy Class	Application	Acceptance Tolerance	Maintenance Tolerance	Special Test Tolerance ¹
0.3	- Petroleum products delivered from large capacity (flow rates greater than 115 L/min or 30 gpm)** devices, including motor-fuel devices - Heated products (other than asphalt) at temperatures greater than 50 °C (122 °F) - Asphalt at temperatures equal to or below 50 °C (122 °F) - All other liquids not shown in the table where the typical delivery is over 200 L (50 gal)	0.2 %	0.3 %	0.5 %
0.3A	- Asphalt at temperatures greater than 50 °C (122 °F)	0.3 %	0.3 %	0.5 %
0.5*	- Petroleum products delivered from small capacity (at 4 L/min (1 gpm) through 115 L/min or 30 gpm)** motor-fuel devices - Agri-chemical liquids - All other applications not shown in the table where the typical delivery is ≤ 200 L (50 gal)	0.3 %	0.5 %	0.5 %
1.1	- Petroleum products and other normal liquids from devices with flow rates** less than 1 gpm. - Devices designed to deliver less than 1 gal	0.75 %	1.0 %	1.25 %
* For test drafts ≤ 40 L or 10 gal, the tolerances specified for Accuracy Class 0.5 in the table above do not apply. For these test drafts, the following applies: (a) Maintenance tolerances on normal and special tests shall be 20 mL plus 4 mL per indicated liter or 1 in³ plus 1 in³ per indicated gallon. (b) Acceptance tolerances on normal and special tests shall be one-half the maintenance tolerance values. ¹ Special test tolerances are not applicable to retail motor fuel and retail Diesel Exhaust Fluid (DEF) dispensers. ** Flow rate refers to designed or marked maximum flow rate.				

(Added 2002) (Amended 2006, 2013, and 2022)

Tolerance Values

Table T.2.

3rd Row in Table -

Applies to small capacity meters or meters that deliver 30 gpm or less will have the following tolerances:

If using a test draft of 10 gallons or less follow the astric * to the bottom of the table

(a) Maintenance tolerances are 1 cubic inch plus 1 cubic inch for every gallon in the test draft

Example: 5-gallon test draft maintenance tolerance is + or – 6 cubic inches

(b) acceptance tolerances are one-half maintenance
Using the above example, acceptance tolerance is + or – 3 cubic inches on 5-gallon test draft



Accuracy Class	Application	Acceptance Tolerance	Maintenance Tolerance	Special Test Tolerance ¹
0.3	- Petroleum products delivered from large capacity (flow rates greater than 115 L/min or 30 gpm)** devices, including motor-fuel devices - Heated products (other than asphalt) at temperatures greater than 50 °C (122 °F) - Asphalt at temperatures equal to or below 50 °C (122 °F) - All other liquids not shown in the table where the typical delivery is over 200 L (50 gal)	0.2 %	0.3 %	0.5 %
0.3A	- Asphalt at temperatures greater than 50 °C (122 °F)	0.3 %	0.3 %	0.5 %
0.5*	- Petroleum products delivered from small capacity (at 4 L/min (1 gpm) through 115 L/min or 30 gpm)** motor-fuel devices - Agri-chemical liquids - All other applications not shown in the table where the typical delivery is ≤ 200 L (50 gal)	0.3 %	0.5 %	0.5 %
1.1	- Petroleum products and other normal liquids from devices with flow rates** less than 1 gpm. - Devices designed to deliver less than 1 gal	0.75 %	1.0 %	1.25 %

* For test drafts ≤ 40 L or 10 gal, the tolerances specified for Accuracy Class 0.5 in the table above do not apply. For these test drafts, the following applies:

(a) Maintenance tolerances on normal and special tests shall be 20 mL plus 4 mL per indicated liter or 1 in³ plus 1 in³ per indicated gallon.

(b) Acceptance tolerances on normal and special tests shall be one-half the maintenance tolerance values.

¹ Special test tolerances are not applicable to retail motor fuel and retail Diesel Exhaust Fluid (DEF) dispensers.

** Flow rate refers to designed or marked maximum flow rate.

Pretest Determinations and Notes

UR.2.5. Product Storage Identification

- The fill Connection for any petroleum product storage tank or vessel shall be permanently, plainly, and visibly marked as to product contained
- When the fill connection is marked by means of a color code, the color code key shall be conspicuously displayed at the place of business



Product Storage: if tanks are marked by color location MUST have a Color Code Posted Inside Business identifying product.



Pretest Determinations and Notes

If the test measure is dry, add 1 in³ to gauge reading to allow for amount of liquid required to “wet” measure

- **N.4.4.1 Pour and Drain Times for Hand-Held Test Measures¹**
 - Handheld Test measures require 30 second (± 5 s) followed by a 10 second drain, with measure held at a 10-15 degree angle from vertical
- **N.4.4.2 Drain Times for Bottom Drain Test Measures or Provers²**
 - Bottom Drain provers require a 30 second drain after main flow ceases



Pretest Determinations and Notes

- Record totalizer indication before each test draft
- After each test print and check receipts for accuracy
- Throughout the test be sure to check for mathematical agreement in computations and between primary and secondary indicators
- Check display of quantity and total price – must display for minimum of five minutes



Pretest Determinations and Notes

Definitions

- Under-registration
 - Meter indicates less product than delivered
 - Buyers favor
 - (+) Reading
- Over-registration
 - Meter indicates more product than delivered
 - Sellers Favor
 - (-) Reading



Pretest Determinations and Notes

Definition continued...

Normal Flow Test

- **N.4.1. Normal Test¹**

- Also called a Full Flow Test
- These tests are conducted at the maximum flow rate of the device (handle on nozzle fully compressed).
Or down to any rate of flow that is half the sum of the max and min flow rate. Within that range.

Special Flow Test

- **N.4.2. Special Test**

- Also called a Slow Flow Test
- Any test except as set forth in N.4.1. Normal Test. Below the range defined in normal test.



Pretest Determinations and Notes

Definition continued...

Special Flow Test shall be conducted as follows:

- On RMFDs **without** a marked minimum flow rate, test conducted at the lesser of the two:
 - At 5 gallons per minute or
 - At the minimum rate when the automatic nozzle is at its lowest setting
- On RMFDs **with** a marked minimum flow rate, test shall be conducted either
 - At or near marked minimum flow rate

What devices are required to have a marked flow rate?





EPO No. 21

Testing Non-Blend Devices

Testing Non-Blend Devices

Non-Blend Devices (Non-Blend Products) will require

- One Normal Test – Full flow test
- One Special Test – Slow flow test

Example - if using 5-gallon test measure, this would require at minimum, 10 gallons of fuel to be dispensed.



Testing Non-Blend Devices



How would you test this product?



Test Non-Blend Devices



Non-Blend Devices (products) will require both a normal (full flow) and special test (slow flow).





EPO No. 22

Testing Blend Devices

Testing Blend Devices

Blend devices (Products) will require testing in this order:

1. A Normal Test & Special Test on the **lowest-grade product** that makes up the blend
2. A Normal Test & Special Test on the **highest-grade product** that makes up the blend
3. A Special Test or Slow flow test on the **blend** product

Example - if using a 5-gallon test measure, at minimum 15 gallons of fuel to be dispensed



Testing Blend Devices



In what order would you test these products and what type of test would you perform on each grade?



Testing Blend Devices



Order: 87 Normal Test & Special Test, 91 Normal Test & Special Test,
89 Special Test



Testing Blend Devices

If a device contains more than one blended product, **after first testing all the non-blended products** test the blends in this order :

1. Test the first blend above the lowest grade
2. Then test the blend below the highest grade

Always return blended product to the storage tank containing the lowest octane




Testing Blend Devices



If the circled grades are the blends, in what order would you test these products, and what type of tests would your perform?



E85 Normal Test & Special Test, 87 Normal Test & Special Test, 91 Normal Test & Special Test, 88 Special Test, 93 Special Test



Testing Blend Devices

Revisions to EPO No. 22 test procedure implemented in 2025

ANY non-blend product will require a normal and special test



Testing Devices Continued...

After test on both non-blend and blend devices, check for the following:

- **S.3.7. Check for effectiveness of anti-drain means**
 - Wet-hose pressure type device
 - Means incorporated to prevent drainage of discharge hose
- **G-UR.4.5. Ensure a Security Seal is affixed**
 - A security seal shall be appropriately affixed to any adjustment mechanism designed to be sealed.



NIST Retail Motor Fuel Dispenser (RMFD) TEST VIDEO

https://www.youtube.com/watch?v=m9HG_3SIsEM



Other Information

**Arkansas Bureau of Standards may soon require
NIST Handbook 44 Certification on for RMFD**

**Certification exam is available on the National
Council on Weights and Measures
(NCWM) website and will be recognized by this
state.**



Summary

- Our goal is to follow up, determine compliance, and exit.
- We perform inspections first and then check pump accuracy.
- Affix Seals and Decals properly.
- Leave a copy of Test Report at the site.
- Email one copy to Bureau@agriculture.arkansas.gov within two weeks of service.
- Once the reports are signed, your agency could be held accountable for non-compliance
- We require your full cooperation when necessary
- Do not work in the State if your Registration has expired.
 - Extension on Registration will not be granted if your Registration has expired
 - If your calibration certification date is over six-months-old, you may not apply for registration with our office
- Your work represents our office and the State of Arkansas.

